

AND

Wrist Digital Blood Pressure Monitor

Model UB-525



Instruction Manual
Original

ENGLISH

Manuel d'instructions
Traduction

FRANÇAIS

Manual de Instrucciones
Traducción

ESPAÑOL

Manuale di Istruzioni
Traduzione

ITALIANO

使用手冊
翻譯

中文

1WMPD4003120E

Contents

| | |
|--|----|
| Dear Customers..... | 2 |
| Preliminary Remarks | 2 |
| Precautions..... | 2 |
| Parts Identification | 4 |
| Symbols..... | 5 |
| Using the Monitor..... | 7 |
| 1. Installing / Changing the Batteries | 7 |
| 2. Applying the Cuff..... | 8 |
| 3. How to Take Accurate Measurements | 8 |
| 4. Measurement..... | 9 |
| 5. After Measurement..... | 9 |
| Measurements..... | 10 |
| Recalling the Memory Data | 11 |
| Deleting Data Stored in Memory..... | 11 |
| What Is The IHB/AFib Indicator? | 12 |
| What Is The AFib? | 12 |
| Pressure Bar Indicator..... | 12 |
| WHO Classification Indicator | 13 |
| About Blood Pressure..... | 13 |
| What is Blood Pressure?..... | 13 |
| What is Hypertension and How is it Controlled? | 13 |
| Why Measure Blood Pressure at Home?..... | 14 |
| WHO Blood Pressure Classification..... | 14 |
| Blood Pressure Variations..... | 14 |
| Troubleshooting..... | 15 |
| Maintenance | 16 |
| Technical Data..... | 16 |

Dear Customers

Congratulations on purchasing a state-of-the-art A&D blood pressure monitor, one of the most advanced monitors available today. Designed for ease of use and accuracy, this device will facilitate your daily blood pressure regimen. **We recommend that you read through this manual carefully before using the device for the first time.**

Preliminary Remarks

- This device conforms to the European Directive 93/42 EEC for Medical Products. This is made evident by the **CE**₀₁₂₃ mark of conformity. (0123: The reference number to the involved notified body)
- The device is designed for use on adults, not newborns or infants.
- Environment for use. The device is for use to operate by yourself in the home healthcare environment.
- This device is designed to measure blood pressure and pulse rate of people for diagnosis.

Precautions

- Precision components are used in the construction of this device. Extremes in temperature, humidity, direct sunlight, shock or dust should be avoided.
- Clean the device with a dry, soft cloth or a cloth dampened with water and a neutral detergent. Never use alcohol, benzene, thinner or other harsh chemicals to clean the device.
- Avoid tightly folding the cuff for long periods, as such treatment may shorten the life of the components.
- The device is not water resistant. Prevent rain, sweat and water from soiling the device.
- Measurements may be distorted if the device is used close to televisions, microwave ovens, cellular telephones, X-ray or other devices with strong electrical fields.
- Used equipment, parts and batteries are not treated as ordinary household waste, and must be disposed of according to the applicable local regulations.
- When reusing the device, confirm that the device is clean.
- Do not modify the device. It may cause accidents or damage to the device.
- To measure blood pressure, the wrist must be squeezed by the cuff hard enough to temporarily stop blood flow through the artery. This may cause pain, numbness or a temporary red mark to the wrist. This condition will appear especially when measurement is repeated successively. Any pain, numbness, or red marks will disappear with time.



- ❑ Wireless communication device, such as home networking devices, mobile phones, cordless phones and their base stations, walkie-talkies can affect this blood pressure monitor. Therefore, a minimum distance of 30 cm should be kept from such devices.
- ❑ Measuring blood pressure too frequently may cause harm due to blood flow interference. Check that the operation of the device does not result in prolonged impairment of blood circulation, when using the device repeatedly.
- ❑ Clinical testing has not been conducted on newborn infants and pregnant woman. Do not use on newborn infants or pregnant woman.
- ❑ If you have had a mastectomy, please consult a doctor before using the device.
- ❑ Do not let children use the device by themselves and do not use the device in a place within the reach of infants. It may cause accidents or damage.
- ❑ There are small parts that may cause a choking hazard if swallowed by mistake by infants.
- ❑ Do not touch the batteries and the patient at the same time. That may result in electrical shock.
- ❑ In the case of single components failure enclosure of near cuff may become hot and potentially cause malfunction.
- ❑ Use of accessories not detailed in this manual may compromise safety.
- ❑ Should the battery short-circuit, it may become hot and potentially cause burns.
- ❑ Allow the device to adapt to the surrounding environment before use (about one hour).
- ❑ Do not inflate without wrapping the cuff around your wrist.

Contraindications

The following are precautions for proper use of the device.


- ❑ Do not apply the device to a wrist with other medical electrical equipment attached. The equipment may not function properly.
- ❑ People who have a severe circulatory deficit in the arm must consult a doctor before using the device to avoid medical problems.
- ❑ Do not self-diagnose the measurement results and start treatment by yourself. Always consult your doctor for evaluation of the results and treatment.
- ❑ Do not apply the device on a wrist with an unhealed wound.
- ❑ Do not apply the device on an arm receiving an intravenous drip or blood transfusion. It may cause injury or accidents.
- ❑ Do not use the device where flammable gases such as anesthetic gases are present. It may cause an explosion.
- ❑ Do not use the device in highly concentrated oxygen environments, such as a high-pressure oxygen chamber or an oxygen tent. It may cause a fire or explosion.

Parts Identification



Symbols

Symbols that are printed on the device case

| Symbols | Function / Meaning |
|--|--|
|  | Standby and Turn the device on |
| SYS | Systolic blood pressure in mmHg |
| DIA | Diastolic blood pressure in mmHg |
| PUL | Pulse per minute |
|  | Battery installation guide |
|  | Direct current |
| SN | Serial number |
| 2015  | Date of manufacture |
|  | Type BF: Device and cuff are designed to provide special protection against electrical shocks. |
|  0123 | EC directive medical device label |
| IP | International protection symbol |
|  | WEEE label |
|  | Manufacturer |
|  | EU-representative |
|  | Refer to instruction manual / booklet |
|  | Keep dry |

Symbols that appear on the display

| Symbols | Function / Meaning / Recommended Action |
|---|--|
|  | Appears while measurement is in progress. It blinks when the pulse is detected. Remain as still as possible. |
|  | IHB/AFib symbol appears when an irregular heartbeat is detected. It may light when a very slight vibration like shivering or shaking is detected. |

Symbols that appear on the display (continued)

| Symbols | Function / Meaning | Recommended Action |
|---|---|---|
|  | Previous measurements stored in MEMORY | _____ |
|  | Average data | _____ |
|  | FULL BATTERY The battery power indicator during measurement | _____ |
|  | LOW BATTERY The battery is low when it blinks | Replace all batteries with new ones when the indicator blinks. |
| E_1 or E_2 | Unstable blood pressure due to movement during measurement | Take another measurement. Remain still during measurement. |
| | The systolic and diastolic values are within 10 mmHg of each other. | |
| E_3 | The pressure value did not increase during inflation. The cuff is not applied correctly. | Apply the cuff correctly, and take another measurement. |
| E | PUL.DISPLAY ERROR The pulse is not detected correctly. | |
| E_E | Blood pressure monitor internal error | Remove the batteries and press the START button, and then install the batteries again. If the error still appears, contact the dealer. |
| E_g | | |


Using the Monitor

1. Installing / Changing the Batteries

1. Remove the battery cover.
2. Remove the used batteries and insert new batteries into the battery compartment as shown, taking care that the polarities (+ and -) are correct. Use only LR03 or AAA batteries.
3. Attach the battery cover.

⚠ CAUTIONS


- ❑ Insert the batteries as shown in the battery compartment. If installed incorrectly, the device will not work.
- ❑ When  (LOW BATTERY mark) blinks on the display, replace all batteries with new ones. Do not mix old and new batteries. It may shorten the battery life, or cause the device to malfunction.
- ❑  (LOW BATTERY mark) does not appear when the batteries are drained.
- ❑ The battery life varies with the ambient temperature and may be shorter at low temperatures. Generally, two new LR03/AAA batteries will last approximately for four months when used twice for measurement each day.
- ❑ Use the specified batteries only. The batteries provided with the device are for testing monitor performance and may have a limited life.
- ❑ Remove the batteries if the device is not to be used for a long period of time. The batteries may leak and cause a malfunction.



2. Applying the Cuff

1. Wrap the cuff around your wrist about 1 cm above your hand as shown in the figure at the right.
2. Apply the cuff tightly using the Velcro strip.

Note: For accurate measurements, apply the cuff tightly and measure on a bare wrist.



3. How to Take Accurate Measurements

For the most accurate blood pressure measurement:

- Remain still and keep quiet during measurement.
- Sit down in a comfortable position. Place your elbow on a table with your palm facing upward and the cuff is at the same level as your heart.
- Relax for about five to ten minutes before measurement. If you are excited or depressed by emotional stress, the measurement will reflect this stress as a higher (or lower) than normal blood pressure reading and the pulse reading will usually be faster than normal.
- Try to measure your blood pressure at about the same time every day.
- An individual's blood pressure varies constantly, depending on what they are doing, what they have eaten and what they drink can have a very strong and rapid effect on your blood pressure.
- Do not measure immediately after physical exercise or a bath. Rest for twenty or thirty minutes before taking the measurement.
- Do not cross your legs. Keep your feet flat on the floor and straighten your back.
- This device bases its measurements on the heartbeat. If you have a very weak or irregular heartbeat, the device may have difficulty determining your blood pressure.
- Should the device detect a condition that is abnormal, it will stop the measurement and display an error symbol. See page 6 for the description of symbols.
- This blood pressure monitor is intended for use by adults. Consult with your physician before using this device on a child. A child should not use this device unattended.
- The automatic blood pressure monitor's performance may be affected by excessive temperature or humidity, or altitude.

4. Measurement

During measurement, it is normal for the cuff to feel very tight.

5. After Measurement

While readings are displayed, if you press the START button to turn the device off, new readings are stored in memory.

While readings are displayed, if you press the ◀ button to turn the device off, new readings are not stored. Remove the cuff and record your data.

Notes: The device is provided with an automatic power shut-off function which stores the current data in the memory and turns the device off automatically one minute after measurement. Allow at least three minutes between measurements on the same person.

Measurements

Note: The UB-525, once used, will provide an inflation appropriate to the user.

1. Wrap the cuff around your wrist. Sit comfortably with the cuff at the same level as your heart and relax.
2. Press the **START** button. All of the display segments are displayed.
3. Zero (0) is displayed blinking briefly. Then the display changes, as measurement begins. The cuff starts to inflate. It is normal for the cuff to feel very tight. The measurement starts automatically when inflation starts, and the ♥ (heart mark) blinks.


Note: If you wish to stop inflation at any time, press the **START** button again.

4. When the measurement is complete, the device displays readings (of the systolic and diastolic pressure readings, pulse rate, WHO classification, I.H.B. symbol). The cuff exhausts the remaining air and completely deflates automatically.

Note: If you do not want to store new readings in memory, press the ◀ button while readings are displayed.

5. Press the **START** button again to turn the device off. Remove the cuff.


Notes: The device is provided with an automatic power shut-off function.
Allow at least three minutes between measurements on the same person.



Recalling the Memory Data


Note: This device stores the last 60 measurements in memory.

- Press the ◀ button when turning the device off.
The average of all measurements and the number of data are displayed.
If no data, "0" is displayed. Press the ◀ or **START** button to turn the device off.
- Use the following buttons to display data (of number and measurement data).
 - The device displays the average of all measurements and the number of data are displayed.
 - Pressing the ◀ button each time, the device displays as follows:
 - Average data of all measurements taken between.
 - Data (of number and measurement data).
 The device displays in order from most recent data.
In the example: No.30 & data → No.29 & data → ... → No.01 & data.
- If you press the ◀ button after oldest data is displayed, the device proceeds to step 1, the average of all measurements and the number of data are displayed.
- Press the **START** button to turn the device off.
After one minute of non-operation, the device will turn off automatically.



Deleting Data Stored in Memory

- Press and hold the ◀ button until turning off the device automatically.
The device displays the **M** mark, deletes data stored in memory while blinking the **M** mark and turns off automatically.



What Is The IHB/AFib Indicator?

When the monitor detects an irregular rhythm during the measurements, the IHB/AFib indicator will appear on the display with the measurement values.


Note: We recommend contacting your physician if you see this «♥» IHB/AFib indicator frequently.

What Is The AFib?

The heart contracts due to electrical signals occurring in heart and sends blood through the body. Atrial fibrillation (AFib) occurs when the electrical signal in the atrium becomes confused and leads to disturbances in the pulse interval. AFib can cause blood to stagnate in the heart, which can easily create clots of blood, a cause of stroke and heart attack.

Pressure Bar Indicator


The indicator monitors the progress of pressure during measurement.




WHO Classification Indicator

Each six segments of the bar indicator correspond to the WHO blood pressure classification is described on page 14.

Example

| Moderate hypertension | Mild hypertension | High normal |
|---|---|--|
|  |  |  |

WHO Classification Indicator



█ : The indicator displays a segment, based on the current data, corresponding to the WHO classification.

About Blood Pressure

What is Blood Pressure?

Blood pressure is the force exerted by blood against the walls of the arteries. Systolic pressure occurs when the heart contracts. Diastolic pressure occurs when the heart expands. Blood pressure is measured in millimeters of mercury (mmHg). One's natural blood pressure is represented by the fundamental pressure, which is measured first thing in the morning while one is still at rest and before eating.

What is Hypertension and How is it Controlled?

Hypertension, an abnormally high arterial blood pressure, if left unattended can cause many health problems including stroke and heart attack. Hypertension can be controlled by altering lifestyle, avoiding stress, and with medication under a doctor's supervision.

To prevent hypertension or keep it under control:

- Do not smoke
- Reduce salt and fat intake
- Maintain proper weight
- Exercise regularly
- Have regular physical checkups

Why Measure Blood Pressure at Home?

Blood pressure measured at a clinic or doctor's office may cause apprehension and can produce an elevated reading, 25 to 30 mmHg higher than that measured at home. Home measurement reduces the effects of outside influences on blood pressure readings, supplements the doctor's readings and provides a more accurate, complete blood pressure history.



WHO Blood Pressure Classification

Standards to assess high blood pressure, without regard to age, have been established by the World Health Organization (WHO), as shown in the chart at the right.

Blood Pressure Variations

An individual's blood pressure varies greatly on a daily and seasonal basis. It may vary by 30 to 50 mmHg due to various conditions during the day. In hypertensive individuals variations are even more pronounced. Normally, the blood pressure rises while at work or play and falls to its lowest levels during sleep. So, do not be overly concerned by the results of one measurement.

Take measurements at the same time every day using the procedure described in this manual to get to know your normal blood pressure. Regular readings give a more comprehensive blood pressure history. Be sure to note the date and time when recording your blood pressure. Consult your doctor to interpret your blood pressure data.



Troubleshooting

| Problem | Possible Reason | Recommended Action |
|--|---|--|
| Nothing appears in the display, even when the device is turned on. | Batteries are drained. | Replace all batteries with new ones. |
| | Battery terminals are not in the correct position. | Reinstall the batteries with negative and positive terminals matching those indicated in the battery compartment. |
| The cuff does not inflate. | Battery voltage is too low.  (LOW BATTERY mark) blinks. If the batteries are drained completely, the mark does not appear. | Replace all batteries with new ones. |
| The device does not measure. Readings are too high or too low. | The cuff is not applied properly. | Apply the cuff correctly. |
| | You moved your wrist or body during the measurement. | Make sure you remain very still and quiet during the measurement. |
| | The cuff position is not correct. | Sit comfortably and still. Place your arm on a table with your palm facing upward and the cuff at the same level as your heart. |
| | ————— | If you have a very weak or irregular heartbeat, the device may have difficulty in determining your blood pressure. |
| Other | The value is different from that measured at a clinic or doctor's office. | See the section "Why Measure Blood Pressure at Home?". |
| | ————— | Remove the batteries. Place them back properly and try measurement again. |

Note: If the actions described above do not solve the problem, contact the dealer. Do not attempt to open or repair this product yourself, as any attempt to do so will make your warranty invalid.

Maintenance

Do not open the device. It uses delicate electrical components and an intricate air unit that could be damaged. If you cannot fix the problem using the troubleshooting instructions, contact the authorized dealer in your area or our customer service department. The A&D customer service will provide technical information, spare parts and units to authorized dealers.

The device was designed and manufactured for a long service life. However it is generally recommended to have the device inspected every 2 years, to ensure proper functioning and accuracy. Please contact the authorized dealer in your area or A&D for maintenance.

Technical Data

| Type | UB-525 |
|------------------------|---|
| Measurement method | Oscillometric measurement |
| Measurement range | Pressure: 0 – 299 mmHg Systolic pressure: 60 – 279 mmHg Diastolic pressure: 40 – 200 mmHg Pulse: 40 – 180 beats / minute |
| Measurement accuracy | Pressure: ± 3 mmHg Pulse: ± 5 % |
| Power supply | 2 x 1.5 V alkaline batteries (LR03 or AAA) |
| Number of measurements | Approx. 250 measurements, when AAA alkaline batteries are used, with pressure value of 170 mmHg at room temperature of 23 °C. |
| Wrist circumference | 13.5 – 21.5 cm |
| Classification | Internally powered ME equipment (Continuous operation mode) |
| Applied part | Cuff Type BF  |
| Useful life | Device: 5 years (when used six times a day) |

| | |
|--------------------------------|--|
| Clinical test | According to ISO81060-2 : 2013 |
| EMC | IEC 60601-1-2: 2014 |
| Memory | Last 60 measurements |
| Operating conditions | +10 to +40 °C / 15 to 85 %RH / 800 to 1060 hPa |
| Transport / storage conditions | -20 to +60 °C / 10 to 95 %RH / 700 to 1060 hPa |
| Dimensions | Approx. 56 [W] x 88 [H] x 21.5 [D] mm |
| Weight | Approx. 90 g, excluding the batteries |
| Ingress protection | IP20 |

Note: Specifications are subject to change for improvement without prior notice.

IP classification is the degrees of protection provided by enclosures in accordance with IEC 60529. This device is protected against solid foreign objects of 12 mm diameter and greater such as a finger. This device is not protected against water.



A&D Company, Limited

1-243 Asahi , Kitamoto-shi, Saitama 364-8585, JAPAN
Telephone: [81] (48) 593-1111 Fax: [81] (48) 593-1119



Emergo Europe B.V.

Prinsessegracht 20, 2514 AP The Hague, The Netherlands
Telephone: [31] (70) 345-8570 Fax: [31] (70) 346-7299

A&D INSTRUMENTS LIMITED

Unit 24/26 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY
United Kingdom
Telephone: [44] (1235) 550420 Fax: [44] (1235) 550485

A&D ENGINEERING, INC.

1756 Automation Parkway, San Jose, California 95131, U.S.A.
Telephone: [1] (408) 263-5333 Fax: [1] (408)263-0119

A&D AUSTRALASIA PTY LTD

32 Dew Street, Thebarton, South Australia 5031, AUSTRALIA
Telephone: [61] (8) 8301-8100 Fax: [61] (8) 8352-7409

ООО A&D RUS

121357, Российская Федерация, г.Москва, ул. Верейская, дом 17
(Business-Center "Vereyskaya Plaza-2" 121357, Russian Federation, Moscow,
Vereyskaya Street 17)
тел.: [7] (495) 937-33-44 факс: [7] (495) 937-55-66

ООО "ЭЙ энд ДИ РУС"

A&D Technology Trading(Shanghai) Co. Ltd 爱安德技研贸易(上海)有限公司

中国 上海市浦东新区浦东大道138号永华大厦21楼A室 邮编200120
(21F Room A, Majesty Building, No.138 Pudong Avenue, Pudong New Area, Shanghai, 200120, China)
电话: [86] (21) 3393-2340 传真: [86] (21) 3393-2347

A&D INSTRUMENTS INDIA PRIVATE LIMITED ऐ&डी इन्स्ट्रूमेंट्स इण्डिया प्रा० लिमिटेड

509, उद्योग विहार , फेस -5, गुडगांव - 122016, हरियाणा , भारत
(509, Udyog Vihar, Phase-V, Gurgaon - 122 016, Haryana, India)
फोन : 91-124-4715555 फैक्स : 91-124-4715599

